

**§ 878.5050 Surgical smoke precipitator.**

(a) *Identification.* A surgical smoke precipitator is a prescription device intended for clearance of the visual field by precipitation of surgical smoke and other aerosolized particulate matter created during laparoscopic surgery.

(b) *Classification.* Class II (special controls). The special controls for this device are:

(1) Adverse tissue reaction must be mitigated through the following:

(i) Chemical characterization and toxicological risk assessment of the treated surgical smoke.

(ii) Demonstration that the elements of the device that may contact the patient are biocompatible.

(2) Electrical safety and electromagnetic compatibility testing must demonstrate that the device performs as intended.

(3) Software verification, validation, and hazard analysis must be performed.

(4) Performance data must demonstrate the sterility of the patient contacting components of the device.

(5) Performance data must support the shelf life of the sterile components of the device by demonstrating continued functionality, sterility, and package integrity over the identified shelf life.

(6) Animal simulated-use testing must demonstrate that the device performs as intended under anticipated conditions of use. The following performance characteristics must be tested:

(i) Device must be demonstrated to be effectively inserted, positioned, and removed from the site of use.

(ii) Device must be demonstrated to precipitate surgical smoke particulates to clear the visual field for laparoscopic surgeries.

(iii) Device must be demonstrated to be non-damaging to the site of use and animal subject.

(7) Labeling must identify the following:

(i) Detailed instructions for use.

(ii) Electrical safety and electromagnetic compatibility information.

(iii) A shelf life.

[83 FR 4143, Jan. 30, 2018]

**Subpart F—Therapeutic Devices**

**§ 878.5070 Air-handling apparatus for a surgical operating room.**

(a) *Identification.* Air-handling apparatus for a surgical operating room is a device intended to produce a directed, nonturbulent flow of air that has been filtered to remove particulate matter and microorganisms to provide an area free of contaminants to reduce the possibility of infection in the patient.

(b) *Classification.* Class II (special controls). The device, when it is an air handling bench apparatus, an air handling room apparatus, or an air handling enclosure apparatus, is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in § 878.9.

[53 FR 23872, June 24, 1988, as amended at 84 FR 71814, Dec. 30, 2019]

**§ 878.5080 Air-handling apparatus accessory.**

(a) *Identification.* An air-handling apparatus accessory is a supplementary device that is intended to be used with an air-handling apparatus for a surgical operating room. This device provides an interface between the components of the device or can be used to switch electrical power. This generic type of device includes fittings, adapters, couplers, remote switches, and footswitches.

(b) *Classification.* Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in § 878.9.

[84 FR 14870, Apr. 12, 2019]

**§ 878.5350 Needle-type epilator.**

(a) *Identification.* A needle-type epilator is a device intended to destroy the dermal papilla of a hair by applying electric current at the tip of a fine needle that has been inserted close to the hair shaft, under the skin, and into the dermal papilla. The electric current may be high-frequency AC current, high-frequency AC combined with DC current, or DC current only.

(b) *Classification.* Class I (general controls). The device is exempt from the premarket notification procedures in

subpart E of part 807 of this chapter, subject to the limitations in § 878.9.

[53 FR 23872, June 24, 1988, as amended at 61 FR 1123, Jan. 16, 1996; 66 FR 38803, July 25, 2001]

#### § 878.5360 Tweezer-type epilator.

(a) *Identification.* The tweezer-type epilator is an electrical device intended to remove hair. The energy provided at the tip of the tweezer used to remove hair may be radio frequency, galvanic (direct current), or a combination of radio frequency and galvanic energy.

(b) *Classification.* Class I (general controls). The device is exempt from premarket notification procedures in subpart E of part 807 of this chapter subject to § 878.9.

[63 FR 57060, Oct. 26, 1998]

#### § 878.5400 Low level laser system for aesthetic use

(a) *Identification.* A Low Level Laser System for Aesthetic Use is a device using low level laser energy for the disruption of adipocyte cells within the fat layer for the release of fat and lipids from these cells for noninvasive aesthetic use.

(b) *Classification.* Class II (special controls). The special control for this device is the FDA guidance document entitled “Guidance for Industry and Food and Drug Administration Staff; Class II Special Controls Guidance Document: Low Level Laser System for Aesthetic Use.” See § 878.1(e) for the availability of this guidance document.

[76 FR 20842, Apr. 14, 2011]

#### § 878.5650 Topical oxygen chamber for extremities.

(a) *Identification.* A topical oxygen chamber for extremities is a device that is intended to surround a patient’s limb and apply humidified oxygen topically at a pressure slightly greater than atmospheric pressure to aid healing of chronic skin ulcers such as bedsores.

(b) *Classification.* Class II (special controls). The special control for this device is FDA’s “Class II Special Controls Guidance: Topical Oxygen Chamber for Extremities.” See § 878.1(e) for

the availability of this guidance document.

[76 FR 22807, Apr. 25, 2011]

#### § 878.5900 Nonpneumatic tourniquet.

(a) *Identification.* A nonpneumatic tourniquet is a device consisting of a strap or tubing intended to be wrapped around a patient’s limb and tightened to reduce circulation.

(b) *Classification.* Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in § 878.9.

[53 FR 23872, June 24, 1988, as amended at 54 FR 13828, Apr. 5, 1989; 59 FR 63010, Dec. 7, 1994; 66 FR 38803, July 25, 2001]

#### § 878.5910 Pneumatic tourniquet.

(a) *Identification.* A pneumatic tourniquet is an air-powered device consisting of a pressure-regulating unit, connecting tubing, and an inflatable cuff. The cuff is intended to be wrapped around a patient’s limb and inflated to reduce or totally occlude circulation during surgery.

(b) *Classification.* Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in § 878.9.

[53 FR 23872, June 24, 1988, as amended at 61 FR 1123, Jan. 16, 1996; 66 FR 38803, July 25, 2001]

## PART 880—GENERAL HOSPITAL AND PERSONAL USE DEVICES

### Subpart A—General Provisions

Sec.

880.1 Scope.

880.3 Effective dates of requirement for premarket approval.

880.9 Limitations of exemptions from section 510(k) of the Federal Food, Drug, and Cosmetic Act (the act).

### Subpart B [Reserved]

### Subpart C—General Hospital and Personal Use Monitoring Devices

880.2200 Liquid crystal forehead temperature strip.

880.2400 Bed-patient monitor.

880.2420 Electronic monitor for gravity flow infusion systems.